

M400 Electronic Current Burden



Electronic Current Burden M400 is a programmable device used as a burden in the secondary circuits of Current Instrument Transformers with secondary current 1 A during their verification and calibration

FEATURES AND BENEFITS

- Rated current 1 A
- Rated burden power range:
 0,5 to 50 V·A, cos φ = 0,8 at 50 Hz
 0; 0,5 to 50 V·A, cos φ = 1 at 50 Hz
- Cable, contact and internal test set (Comparator) resistances are compensated by four-conductor connection
- Non-breaking measuring circuit during switching of the burden
- Possibility of using with automated CT calibration systems
- Small size and low weight

VERSIONS

Electronic Current Burden M400 is available in two versions:

M400.1 allows CT calibration in secondary current range from 1 % to 120 % of rated current

M400.2 allows CT calibration in secondary current range from 1 % to 150 % of rated current

APPLICATIONS

M400 is used by:

- Manufacturer of Current Transformers
- On-site testing of Current Transformers
- Calibration Laboratories
- Metrology Institutes



TECHNICAL SPECIFICATIONS

Rated current	1 A	M400.1, M400.2	
Command was as	1120 %	M400.1	
Current range	1150 %	M400.2	
Rated burden power for $\cos \phi = 0.8$	0,5; 0,75; 1; 1,25; 1,5; 1,75; 1,875; 2; 2,5; 3; 3,75; 4; 5; 6,25; 7,5; 10; 11,25; 12,5; 15; 20; 25; 30; 40; 45*; 50* V·A	M400.1, M400.2	
Rated burden power for $\cos \varphi = 1$	0; 0,5; 0,75; 0,8; 1; 1,25; 1,5; 1,875; 2; 2,5; 3,75; 4; 5; 6; 7,5; 10; 12,5; 15; 30; 50* V·A	M400.1, M400.2	
	± 0,03·S** for 0,8 V·A < S < 50 V·A		
Active and reactive power error	± 0,04·S for S = 0,75 V·A	M400.1,	
	± 0,06·S for S = 0,5 V·A		
	± 0,05 V·A for S = 0 V·A		

^{*} Burden power values are possible for 1 to 120 % of rated current ** Set burden power value

Power	mains
Dota	d Valtaga

Rated Voltage	220/230 V	
Rated Frequency	50 Hz	
Operating Temperature	040 °C	
Relative Humidity	up to 80 % non-condensing	
Size (Main Unit)	250 × 148 × 330 mm	
Weight (Main Unit)	5 kg	
Standards		
Safety	EN 61010-1:2010	
EMC	EN 61326-1:2013	
Calibration Interval	2 years recommended	



ORDERING INFORMATION

Scope of supply

The following items are supplied with the standard M400 modification:

No	Item Name	Part no.	
1*	Main unit M400.1 M400.2	411648.001-01 411648.001-02	M400 Assertions Command Annexes.
2	IC Instrument cable (1,8 m)	685611.195-01	
3	Serial port interface cable RS232 Null Modem Cable DB9 F/F	_	
4	Power cable 10 A EU (CEE 7/XVII - C13)	_	
5	Jumper wire	685611.030	
6	Bag	323382.007	AN
7	Operating manual	411640.001 OM	
8	Passport	411640.001 P	
	Additional order		
9	IC Instrument cable (5 m)	685611.196-01	
10	IC Instrument cable (10 m)	685611.197-01	

^{*} Please specify modification when ordering

Development and production of measurement devices

OLTEST LLC, PO box 33, Kyiv, 04128, Ukraine

TEL and E-MAIL

+38 044 537-08-01 market@oltest.ua